

Paula Vandervoort

From: John Humphrey <john@medata.us>
Sent: Thursday, September 06, 2012 7:00 PM
To: Paula Vandervoort
Subject: Scans.

Hi Paula,

Ginger is an excellent scan with only one false positive in 96,290 detected beats. Negligable false positive rate which is remarkable.

Scotty had 76 false positives in 86,840 detected beats, a false positive rate of 0.0817%. The reason is that our algorithm has difficulties where there are high T waves in both channels especially when RR variability is pronounced.

Your electrode application is very good because electrode/movement artifact is minimal.

Both dogs look healthy to me with no PVCs detected based on what I see in the ECG strips.

Best regards,

John.